Accredited entity according to ČSN EN ISO/IEC 17025:2018:

MSP - GLOBAL, s.r.o.,
CAB number 2286, Czech Calibration Service
Dobrá 240, 739 51 Dobrá

CMC for the field of measured quantity: Mass

Ordinal number[1]	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand		Lowest stated expanded measurement uncertainty[2]	Calibration principle	Calibration procedure identification[3]	Work-place
		min.	unit	max.	unit						
1*	Electronic and mechanical scales with non-automatic function	0 kg	up to	0.22 kg		class F1	weight	1.4 mg	Direct loading with a reference weight	KBP 14.1, KBP 14.2	
		0.22 kg	up to	1.6 kg		class F1	weight	5.8 mg			
		1.6 kg	up to	5 kg		class F1	weight	42 mg			
		0.5 kg	up to	10 kg		class F1	weight	81 mg			
		10 kg	up to	16 kg		class F2	weight	0.13 g			
		16 kg	up to	20 kg		class F2	weight	1.2 g			
		20 kg	up to	50 kg		class M1	weight	3.2 g			
		50 kg	up to	100 kg		class M1	weight	6.3 g			

[1] Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

[2] The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

[3] If the document identifying the test procedure is dated, only these specific procedures are used. If the document identifying the test procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Český institut pro akreditaci, o.p.s.
NÁRODNÍ AKREDITAČNÍ ORGÁN
-3-